

6/6/2018
Mr. Gustan Taylor
Arcadis U.S., Inc.
28550 Cabot Dr.
Suite 500
Novi MI 48377

Project Name: FORD LTP
Project #: MI001454.0004.00003
Workorder #: 1806042

Dear Mr. Gustan Taylor

The following report includes the data for the above referenced project for sample(s) received on 6/1/2018 at Air Toxics Ltd.

The data and associated QC analyzed by TO-15 are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics Inc. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Ausha Scott at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Ausha Scott
Project Manager

WORK ORDER #: 1806042

Work Order Summary

CLIENT:	Mr. Gustan Taylor Arcadis U.S., Inc. 28550 Cabot Dr. Suite 500 Novi, MI 48377	BILL TO:	Accounts Payable Arcadis U.S., Inc. 630 Plaza Drive Suite 600 Highlands Ranch, CO 80129
PHONE:	248.994.2294	P.O. #	MI001386.0001_AT_WA08
FAX:		PROJECT #	MI001454.0004.00003 FORD LTP
DATE RECEIVED:	06/01/2018	CONTACT:	Ausha Scott
DATE COMPLETED:	06/06/2018		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	SVMP-25D-6(052918)	TO-15	6.1 "Hg	15.6 psi
02A	SVMP-26-4(052918)	TO-15	5.7 "Hg	14.9 psi
03A	SVMP-25S-3(052918)	TO-15	5.5 "Hg	15.3 psi
04A	SVMP-27-4.5(052918)	TO-15	6.1 "Hg	14.5 psi
05A	SVMP-28-3(052918)	TO-15	7.1 "Hg	14.8 psi
06A	SVMP-12-3.5(052918)	TO-15	5.3 "Hg	14.7 psi
07A	SVMP-08-3.5(052918)	TO-15	6.3 "Hg	14.7 psi
08A	SVMP-09-4(053018)	TO-15	4.1 "Hg	14.2 psi
09A	SVMP-07-3.5(053018)	TO-15	5.5 "Hg	15.1 psi
10A	DUP-01(052918)	TO-15	5.9 "Hg	14.4 psi
11A	SVMP-06-4.5(053018)	TO-15	4.3 "Hg	15.3 psi
12A	SVMP-05-4.5(053018)	TO-15	4.7 "Hg	14.5 psi
13A	SVMP-04-3.5(053018)	TO-15	5.5 "Hg	14.3 psi
14A	SVMP-17-2(053018)	TO-15	5.3 "Hg	14.4 psi
15A	DUP-02(053018)	TO-15	5.5 "Hg	14.9 psi
16A	Lab Blank	TO-15	NA	NA
17A	CCV	TO-15	NA	NA
18A	LCS	TO-15	NA	NA
18AA	LCSD	TO-15	NA	NA

CERTIFIED BY: 
 Technical Director

DATE: 06/06/18

Certification numbers: AZ Licensure AZ0775, NJ NELAP - CA016, NY NELAP - 11291,
 TX NELAP - T104704434-16-11, UT NELAP CA0093332016-7, VA NELAP - 8113, WA NELAP - C935
 Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)
 Accreditation number: CA300005, Effective date: 10/18/2016, Expiration date: 10/17/2017.

Eurofins Air Toxics Inc.. certifies that the test results contained in this report meet all requirements of the NELAC standards

This report shall not be reproduced, except in full, without the written approval of Eurofins Air Toxics, Inc.

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630
 (916) 985-1000 . (800) 985-5955 . FAX (916) 985-1020

LABORATORY NARRATIVE
EPA Method TO-15
Arcadis U.S., Inc.
Workorder# 1806042

Fifteen 1 Liter Summa Canister samples were received on June 01, 2018. The laboratory performed analysis via EPA Method TO-15 using GC/MS in the full scan mode.

This workorder was independently validated prior to submittal using 'USEPA National Functional Guidelines' as generally applied to the analysis of volatile organic compounds in air. A rules-based, logic driven, independent validation engine was employed to assess completeness, evaluate pass/fail of relevant project quality control requirements and verification of all quantified amounts.

Receiving Notes

The Chain of Custody (COC) information for sample SVMP-25S-3(052918) did not match the information on the canister with regard to canister identification. The client was notified of the discrepancy and the information on the canister was used to process and report the sample.

Sample collection date was not provided on the Chain of Custody (COC) for samples DUP-02(053018). The sampling date was taken from the tag.

Analytical Notes

There were no analytical discrepancies.

Definition of Data Qualifying Flags

Ten qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

E - Exceeds instrument calibration range.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

M - Reported value may be biased due to apparent matrix interferences.

CN - See Case Narrative.

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

Summary of Detected Compounds EPA METHOD TO-15 GC/MS FULL SCAN

Client Sample ID: SVMP-25D-6(052918)

Lab ID#: 1806042-01A

No Detections Were Found.

Client Sample ID: SVMP-26-4(052918)

Lab ID#: 1806042-02A

No Detections Were Found.

Client Sample ID: SVMP-25S-3(052918)

Lab ID#: 1806042-03A

No Detections Were Found.

Client Sample ID: SVMP-27-4.5(052918)

Lab ID#: 1806042-04A

No Detections Were Found.

Client Sample ID: SVMP-28-3(052918)

Lab ID#: 1806042-05A

No Detections Were Found.

Client Sample ID: SVMP-12-3.5(052918)

Lab ID#: 1806042-06A

No Detections Were Found.

Client Sample ID: SVMP-08-3.5(052918)

Lab ID#: 1806042-07A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Trichloroethene	1.3	2.0	6.8	11

Client Sample ID: SVMP-09-4(053018)

Lab ID#: 1806042-08A

No Detections Were Found.

**Summary of Detected Compounds
EPA METHOD TO-15 GC/MS FULL SCAN**

Client Sample ID: SVMP-07-3.5(053018)

Lab ID#: 1806042-09A

No Detections Were Found.

Client Sample ID: DUP-01(052918)

Lab ID#: 1806042-10A

No Detections Were Found.

Client Sample ID: SVMP-06-4.5(053018)

Lab ID#: 1806042-11A

No Detections Were Found.

Client Sample ID: SVMP-05-4.5(053018)

Lab ID#: 1806042-12A

No Detections Were Found.

Client Sample ID: SVMP-04-3.5(053018)

Lab ID#: 1806042-13A

No Detections Were Found.

Client Sample ID: SVMP-17-2(053018)

Lab ID#: 1806042-14A

No Detections Were Found.

Client Sample ID: DUP-02(053018)

Lab ID#: 1806042-15A

No Detections Were Found.



Air Toxics

Client Sample ID: SVMP-25D-6(052918)

Lab ID#: 1806042-01A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060512	Date of Collection:	5/29/18 11:13:00 AM
Dil. Factor:	2.59	Date of Analysis:	6/5/18 03:58 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.3	Not Detected	3.3	Not Detected
1,1-Dichloroethene	1.3	Not Detected	5.1	Not Detected
cis-1,2-Dichloroethene	1.3	Not Detected	5.1	Not Detected
Trichloroethene	1.3	Not Detected	7.0	Not Detected
trans-1,2-Dichloroethene	1.3	Not Detected	5.1	Not Detected
Tetrachloroethene	1.3	Not Detected	8.8	Not Detected
1,4-Dioxane	5.2	Not Detected	19	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	100	70-130
1,2-Dichloroethane-d4	93	70-130
4-Bromofluorobenzene	116	70-130



Air Toxics

Client Sample ID: SVMP-26-4(052918)

Lab ID#: 1806042-02A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060513	Date of Collection:	5/29/18 1:15:00 PM
Dil. Factor:	2.48	Date of Analysis:	6/5/18 04:26 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.2	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.9	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Trichloroethene	1.2	Not Detected	6.7	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Tetrachloroethene	1.2	Not Detected	8.4	Not Detected
1,4-Dioxane	5.0	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	101	70-130
1,2-Dichloroethane-d4	94	70-130
4-Bromofluorobenzene	113	70-130



Air Toxics

Client Sample ID: SVMP-25S-3(052918)

Lab ID#: 1806042-03A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060514	Date of Collection:	5/29/18 11:50:00 AM
Dil. Factor:	2.50	Date of Analysis:	6/5/18 04:55 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.2	Not Detected
1,1-Dichloroethene	1.2	Not Detected	5.0	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	5.0	Not Detected
Trichloroethene	1.2	Not Detected	6.7	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	5.0	Not Detected
Tetrachloroethene	1.2	Not Detected	8.5	Not Detected
1,4-Dioxane	5.0	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	103	70-130
1,2-Dichloroethane-d4	95	70-130
4-Bromofluorobenzene	114	70-130



Air Toxics

Client Sample ID: SVMP-27-4.5(052918)

Lab ID#: 1806042-04A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060517	Date of Collection:	5/29/18 2:11:00 PM
Dil. Factor:	2.49	Date of Analysis:	6/5/18 06:20 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.2	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.9	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Trichloroethene	1.2	Not Detected	6.7	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Tetrachloroethene	1.2	Not Detected	8.4	Not Detected
1,4-Dioxane	5.0	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	96	70-130
1,2-Dichloroethane-d4	93	70-130
4-Bromofluorobenzene	126	70-130



Air Toxics

Client Sample ID: SVMP-28-3(052918)

Lab ID#: 1806042-05A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060516	Date of Collection:	5/29/18 3:03:00 PM
Dil. Factor:	2.63	Date of Analysis:	6/5/18 05:51 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.3	Not Detected	3.4	Not Detected
1,1-Dichloroethene	1.3	Not Detected	5.2	Not Detected
cis-1,2-Dichloroethene	1.3	Not Detected	5.2	Not Detected
Trichloroethene	1.3	Not Detected	7.1	Not Detected
trans-1,2-Dichloroethene	1.3	Not Detected	5.2	Not Detected
Tetrachloroethene	1.3	Not Detected	8.9	Not Detected
1,4-Dioxane	5.3	Not Detected	19	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	111	70-130
1,2-Dichloroethane-d4	93	70-130
4-Bromofluorobenzene	112	70-130



Air Toxics

Client Sample ID: SVMP-12-3.5(052918)

Lab ID#: 1806042-06A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060515	Date of Collection:	5/29/18 3:55:00 PM
Dil. Factor:	2.43	Date of Analysis:	6/5/18 05:23 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.1	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.8	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Trichloroethene	1.2	Not Detected	6.5	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Tetrachloroethene	1.2	Not Detected	8.2	Not Detected
1,4-Dioxane	4.9	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	92	70-130
1,2-Dichloroethane-d4	91	70-130
4-Bromofluorobenzene	115	70-130



Air Toxics

Client Sample ID: SVMP-08-3.5(052918)

Lab ID#: 1806042-07A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060518	Date of Collection:	5/29/18 4:36:00 PM
Dil. Factor:	2.53	Date of Analysis:	6/5/18 08:53 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.3	Not Detected	3.2	Not Detected
1,1-Dichloroethene	1.3	Not Detected	5.0	Not Detected
cis-1,2-Dichloroethene	1.3	Not Detected	5.0	Not Detected
Trichloroethene	1.3	2.0	6.8	11
trans-1,2-Dichloroethene	1.3	Not Detected	5.0	Not Detected
Tetrachloroethene	1.3	Not Detected	8.6	Not Detected
1,4-Dioxane	5.1	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	94	70-130
1,2-Dichloroethane-d4	92	70-130
4-Bromofluorobenzene	115	70-130



Air Toxics

Client Sample ID: SVMP-09-4(053018)

Lab ID#: 1806042-08A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060521	Date of Collection:	5/30/18 6:43:00 AM
Dil. Factor:	2.28	Date of Analysis:	6/5/18 10:18 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.1	Not Detected	2.9	Not Detected
1,1-Dichloroethene	1.1	Not Detected	4.5	Not Detected
cis-1,2-Dichloroethene	1.1	Not Detected	4.5	Not Detected
Trichloroethene	1.1	Not Detected	6.1	Not Detected
trans-1,2-Dichloroethene	1.1	Not Detected	4.5	Not Detected
Tetrachloroethene	1.1	Not Detected	7.7	Not Detected
1,4-Dioxane	4.6	Not Detected	16	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	98	70-130
1,2-Dichloroethane-d4	94	70-130
4-Bromofluorobenzene	114	70-130



Air Toxics

Client Sample ID: SVMP-07-3.5(053018)

Lab ID#: 1806042-09A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060522	Date of Collection:	5/30/18 8:03:00 AM
Dil. Factor:	2.48	Date of Analysis:	6/5/18 10:46 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.2	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.9	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Trichloroethene	1.2	Not Detected	6.7	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Tetrachloroethene	1.2	Not Detected	8.4	Not Detected
1,4-Dioxane	5.0	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	98	70-130
1,2-Dichloroethane-d4	94	70-130
4-Bromofluorobenzene	113	70-130



Air Toxics

Client Sample ID: DUP-01(052918)

Lab ID#: 1806042-10A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060519	Date of Collection:	5/29/18
Dil. Factor:	2.46	Date of Analysis:	6/5/18 09:21 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.1	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.9	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Trichloroethene	1.2	Not Detected	6.6	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Tetrachloroethene	1.2	Not Detected	8.3	Not Detected
1,4-Dioxane	4.9	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	98	70-130
1,2-Dichloroethane-d4	92	70-130
4-Bromofluorobenzene	114	70-130



Air Toxics

Client Sample ID: SVMP-06-4.5(053018)

Lab ID#: 1806042-11A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060523	Date of Collection:	5/30/18 9:03:00 AM
Dil. Factor:	2.38	Date of Analysis:	6/5/18 11:14 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.0	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.7	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.7	Not Detected
Trichloroethene	1.2	Not Detected	6.4	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.7	Not Detected
Tetrachloroethene	1.2	Not Detected	8.1	Not Detected
1,4-Dioxane	4.8	Not Detected	17	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	99	70-130
1,2-Dichloroethane-d4	93	70-130
4-Bromofluorobenzene	114	70-130



Air Toxics

Client Sample ID: SVMP-05-4.5(053018)

Lab ID#: 1806042-12A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060524	Date of Collection:	5/30/18 9:53:00 AM
Dil. Factor:	2.36	Date of Analysis:	6/5/18 11:42 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.0	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.7	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.7	Not Detected
Trichloroethene	1.2	Not Detected	6.3	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.7	Not Detected
Tetrachloroethene	1.2	Not Detected	8.0	Not Detected
1,4-Dioxane	4.7	Not Detected	17	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	99	70-130
1,2-Dichloroethane-d4	93	70-130
4-Bromofluorobenzene	113	70-130



Air Toxics

Client Sample ID: SVMP-04-3.5(053018)

Lab ID#: 1806042-13A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060520	Date of Collection:	5/30/18 11:38:00 AM
Dil. Factor:	2.42	Date of Analysis:	6/5/18 09:49 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.1	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.8	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Trichloroethene	1.2	Not Detected	6.5	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Tetrachloroethene	1.2	Not Detected	8.2	Not Detected
1,4-Dioxane	4.8	Not Detected	17	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	98	70-130
1,2-Dichloroethane-d4	92	70-130
4-Bromofluorobenzene	114	70-130



Air Toxics

Client Sample ID: SVMP-17-2(053018)

Lab ID#: 1806042-14A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060525	Date of Collection:	5/30/18 2:28:00 PM
Dil. Factor:	2.40	Date of Analysis:	6/6/18 12:11 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.1	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.8	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Trichloroethene	1.2	Not Detected	6.4	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.8	Not Detected
Tetrachloroethene	1.2	Not Detected	8.1	Not Detected
1,4-Dioxane	4.8	Not Detected	17	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	99	70-130
1,2-Dichloroethane-d4	89	70-130
4-Bromofluorobenzene	117	70-130



Air Toxics

Client Sample ID: DUP-02(053018)

Lab ID#: 1806042-15A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060526	Date of Collection:	5/30/18
Dil. Factor:	2.46	Date of Analysis:	6/6/18 12:39 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	1.2	Not Detected	3.1	Not Detected
1,1-Dichloroethene	1.2	Not Detected	4.9	Not Detected
cis-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Trichloroethene	1.2	Not Detected	6.6	Not Detected
trans-1,2-Dichloroethene	1.2	Not Detected	4.9	Not Detected
Tetrachloroethene	1.2	Not Detected	8.3	Not Detected
1,4-Dioxane	4.9	Not Detected	18	Not Detected

Container Type: 1 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	100	70-130
1,2-Dichloroethane-d4	97	70-130
4-Bromofluorobenzene	110	70-130



Air Toxics

Client Sample ID: Lab Blank

Lab ID#: 1806042-16A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060511	Date of Collection:	NA
Dil. Factor:	1.00	Date of Analysis:	6/5/18 02:43 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Vinyl Chloride	0.50	Not Detected	1.3	Not Detected
1,1-Dichloroethene	0.50	Not Detected	2.0	Not Detected
cis-1,2-Dichloroethene	0.50	Not Detected	2.0	Not Detected
Trichloroethene	0.50	Not Detected	2.7	Not Detected
trans-1,2-Dichloroethene	0.50	Not Detected	2.0	Not Detected
Tetrachloroethene	0.50	Not Detected	3.4	Not Detected
1,4-Dioxane	2.0	Not Detected	7.2	Not Detected

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	100	70-130
1,2-Dichloroethane-d4	91	70-130
4-Bromofluorobenzene	110	70-130

Client Sample ID: CCV

Lab ID#: 1806042-17A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060508	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 6/5/18 01:21 PM

Compound	%Recovery
Vinyl Chloride	74
1,1-Dichloroethene	91
cis-1,2-Dichloroethene	98
Trichloroethene	97
trans-1,2-Dichloroethene	96
Tetrachloroethene	110
1,4-Dioxane	95

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	105	70-130
1,2-Dichloroethane-d4	96	70-130
4-Bromofluorobenzene	113	70-130

Client Sample ID: LCS

Lab ID#: 1806042-18A

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060509	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 6/5/18 01:48 PM

Compound	%Recovery	Method Limits
Vinyl Chloride	77	70-130
1,1-Dichloroethene	91	70-130
cis-1,2-Dichloroethene	88	70-130
Trichloroethene	102	70-130
trans-1,2-Dichloroethene	104	70-130
Tetrachloroethene	110	70-130
1,4-Dioxane	106	70-130

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	98	70-130
1,2-Dichloroethane-d4	94	70-130
4-Bromofluorobenzene	113	70-130

Client Sample ID: LCSD

Lab ID#: 1806042-18AA

EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	17060510	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 6/5/18 02:15 PM

Compound	%Recovery	Method Limits
Vinyl Chloride	77	70-130
1,1-Dichloroethene	92	70-130
cis-1,2-Dichloroethene	86	70-130
Trichloroethene	98	70-130
trans-1,2-Dichloroethene	102	70-130
Tetrachloroethene	114	70-130
1,4-Dioxane	101	70-130

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	102	70-130
1,2-Dichloroethane-d4	97	70-130
4-Bromofluorobenzene	114	70-130